

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	8.2	
4 - 1	7.5	
7 - 1	7.71	
8 - 1	8	
9 - 1	8.2	
10 - 1	8.88	
11 - 1	8.25	
12 - 1	7.74	
14 - 1	7.8	
15 - 1	7.86	
17 - 1	8	
18 - 1	8.02	
21 - 1	7.74	
23 - 1	8.26	
24 - 1	8.47	
25 - 1	8.72	
26 - 1	8.65	
28 - 1	8.63	
30 - 1	7.89	

Westgard	تفسیر	Folic acid	آزمایش	Maglumi	دستگاه	11	ماه	1402	سال
Control	Snibe	Mean	8.16	SD	0.39	Exp. Date	1402/11/25	Lot. No	066220411
Next Month	Mean	8.13	SD	0.39	CV	4.81			

