

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal		High		Low	
Mean	48.83	Mean	58.2	Mean	20.6
SD	0.29	SD	0.31	SD	0.22
CV	0.6	CV	0.53	CV	1.06
Day-Run	Value	Day-Run	Value	Day-Run	Value
2 - 1	49	2 - 1	58.8	2 - 1	20.5
3 - 1	49	3 - 1	57.9	3 - 1	20.6
4 - 1	48.5	4 - 1	58	4 - 1	20.6
5 - 1	49	5 - 1	58	5 - 1	20.6
6 - 1	49	6 - 1	58.3	6 - 1	20.8
9 - 1	49	9 - 1	57.8	9 - 1	21
11 - 1	48.8	11 - 1	58.3	11 - 1	20.6
12 - 1	49	12 - 1	57.8	12 - 1	20.5
19 - 1	48.2	19 - 1	58.3	19 - 1	20.2

Westgard	تفسیر	HCT	آزمایش	NIHON KOHDEN سل کانترا	دستگاه	10	ماه	1402	سال
Normal Control		Mean	48.6	SD	0.46	Exp. Date	1402/12/29	Lot. No	
High Control		Mean	58.2	SD	0.22	Exp. Date	1402/12/29	Lot. No	
Low Control		Mean	20.5	SD	0.25	Exp. Date	1402/12/29	Lot. No	

