

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	143.1	
2 - 2	140.5	
3 - 1	146.2	
3 - 2	141.7	
7 - 1	139.5	
8 - 1	143.6	
9 - 1	140.2	
10 - 1	146.3	
12 - 1	140.3	
14 - 1	142.2	
15 - 1	145.6	
16 - 1	141.1	
17 - 1	143.2	
18 - 1	145.6	
20 - 1	143.5	
23 - 1	138.3	
24 - 1	142.7	
25 - 1	139.7	
28 - 1	143.2	
29 - 1	140.3	
30 - 1	143.2	

Westgard	تفسیر	Ig A	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	141.45	SD	2.6	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	142.38	SD	2.32	CV	1.63			

