

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	775.3	
2 - 1	767.3	
3 - 1	775.3	
4 - 1	770.6	
7 - 1	776.3	
8 - 1	768.9	
9 - 1	770.6	
10 - 1	778	
12 - 1	770.1	
14 - 1	776.2	
15 - 1	768.3	
16 - 1	767	
17 - 1	773	
18 - 1	769.3	
21 - 1	771.4	
23 - 1	775	
24 - 1	771.3	
25 - 1	773	
28 - 1	775	
29 - 1	768.4	
30 - 1	768.1	

Westgard	تفسیر	Ig G	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	771.06	SD	4.66	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	771.83	SD	3.37	CV	0.44			

