

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	49.2	
3 - 1	53.1	
4 - 1	50.6	
5 - 1	48.2	
6 - 1	50.2	
7 - 1	52.5	
8 - 1	55.2	
9 - 1	49.3	
10 - 1	52.6	
11 - 1	55	
12 - 1	51.1	
13 - 1	47.9	
14 - 1	48.1	
16 - 1	48.4	
17 - 1	53.2	
18 - 1	55.2	
19 - 1	54.9	
20 - 1	47.3	
21 - 1	55	
23 - 1	53	
24 - 1	55.2	
25 - 1	51	
26 - 1	51.7	
27 - 1	52	
28 - 1	53	
30 - 1	53.2	

Westgard	تفسیر	Ig M	آزمایش	BS-480 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	51.73	SD	2.72	Exp. Date	1402/11/7	Lot. No	641-01-030/2
Next Month	Mean	51.77	SD	2.58	CV	4.98			

