

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	4.8	
2 - 1	4.7	
3 - 1	4.8	
4 - 1	4.72	
7 - 1	4.79	
8 - 1	4.8	
9 - 1	4.73	
10 - 1	4.86	
12 - 1	4.7	
14 - 1	4.83	
15 - 1	4.8	
16 - 1	4.8	
17 - 1	4.9	
18 - 1	4.7	
21 - 1	4.72	
23 - 1	4.86	
24 - 1	4.8	
25 - 1	4.8	
28 - 1	4.7	
29 - 1	4.8	
30 - 1	4.83	

Westgard	تفسیر	Potassium (K)	آزمایش	AUDICOM	دستگاه	11	ماه	1402	سال
Control	زیست شیمی	Mean	4.76	SD	0.08	Exp. Date	1404/1/14	Lot. No	210202-17
Next Month	Mean	4.78	SD	0.06	CV	1.24			

