

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 2 - 1 | 140 | |
| 3 - 1 | 138 | |
| 4 - 1 | 138 | |
| 5 - 1 | 133 | 1:2S |
| 6 - 1 | 139 | 1:3S |
| 7 - 1 | 135 | |
| 9 - 1 | 135.5 | |
| 10 - 1 | 139.6 | |
| 11 - 1 | 133 | |
| 12 - 1 | 136.8 | |
| 13 - 1 | 142 | |
| 14 - 1 | 136.9 | |
| 16 - 1 | 134 | |
| 17 - 1 | 132.5 | |
| 18 - 1 | 134.5 | |
| 19 - 1 | 139.1 | |
| 20 - 1 | 137.5 | |
| 21 - 1 | 132.8 | |
| 23 - 1 | 135 | |
| 24 - 1 | 139.2 | |
| 25 - 1 | 140 | |
| 26 - 1 | 139 | |
| 27 - 1 | 136.1 | |
| 28 - 1 | 140 | |
| 30 - 1 | 142 | |

| | | | | | | | | | |
|------------|--------|--------|--------|--------------------|--------|-----------|-----------|---------|-----|
| Westgard | تفسیر | LDL | آزمایش | اتوآنالایزر BS-800 | دستگاه | 10 | ماه | 1402 | سال |
| Control | biorex | Mean | 137.36 | SD | 2.76 | Exp. Date | 1403/1/17 | Lot. No | 035 |
| Next Month | Mean | 137.14 | SD | 2.88 | CV | 2.1 | | | |

