

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	32.4	
3 - 1	33.5	
4 - 1	31.9	
5 - 1	30.8	
6 - 1	32.9	
7 - 1	32.9	
9 - 1	30.9	
10 - 1	32.1	
11 - 1	32.6	
12 - 1	30.8	
13 - 1	30.4	
14 - 1	31.1	
16 - 1	32.5	
17 - 1	32.9	
18 - 1	33.1	
19 - 1	31.5	
20 - 1	31.9	
21 - 1	32.4	
23 - 1	30.6	
24 - 1	30.9	
25 - 1	29.7	
26 - 1	30.8	
27 - 1	32.2	
28 - 1	32.9	
30 - 1	31.1	

Westgard	تفسیر	LH	آزمایش	2000 ایمولایت XPI-	دستگاه	10	ماه	1402	سال
Control	Seronorm	Mean	31.32	SD	1.96	Exp. Date	1403/5/31	Lot. No	2011909
Next Month	Mean	31.79	SD	1.02	CV	3.21			

