

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	44.1	
3 - 1	44.5	
4 - 1	42.9	
5 - 1	43.9	
6 - 1	43.2	
7 - 1	44.6	
9 - 1	42.5	
10 - 1	44.1	
11 - 1	43.2	1:2S
12 - 1	42.6	
13 - 1	44.3	
14 - 1	43.3	
16 - 1	44.7	
17 - 1	42.6	
18 - 1	44.2	
19 - 1	42.6	
20 - 1	43.2	
21 - 1	44.7	
23 - 1	42.9	
24 - 1	44	
25 - 1	43.53	
26 - 1	42.9	
27 - 1	43.1	
28 - 1	44.5	
30 - 1	43.6	

Westgard	تفسیر	Lactate(Lactic Acid)	آزمایش	BS-480 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	43.56	SD	0.6	Exp. Date	1402/12/14	Lot. No	
Next Month	Mean	43.59	SD	0.74	CV	1.7			

