

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	45	
2 - 2	42.3	
3 - 1	44.5	
4 - 1	42.7	
7 - 1	44.5	
8 - 1	42.6	
9 - 1	43.2	
10 - 1	44.6	
12 - 1	43	
14 - 1	43.2	
15 - 1	44.3	
16 - 1	43.2	
17 - 1	44.7	
18 - 1	43.2	
21 - 1	44	
23 - 1	42.5	
24 - 1	44.3	
25 - 1	43.2	
28 - 1	42.3	
29 - 1	44.2	
30 - 1	44.2	

Westgard	تفسیر	Lactate(Lactic Acid)	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	43.59	SD	0.74	Exp. Date	1402/12/14	Lot. No	
Next Month	Mean	43.6	SD	0.87	CV	2			

