

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	68	
3 - 1	66	
4 - 1	68	
5 - 1	65.4	
6 - 1	64	
7 - 1	68	
9 - 1	67	
10 - 1	67.2	
11 - 1	68.5	
12 - 1	64.8	
13 - 1	65.7	
14 - 1	67.4	
16 - 1	65.3	
17 - 1	67.9	
18 - 1	64	
19 - 1	68	
20 - 1	70	
21 - 1	65.6	
23 - 1	68	
24 - 1	65.2	
25 - 1	68	
26 - 1	66.2	
27 - 1	68.5	
28 - 1	66	
30 - 1	66.1	

Westgard	تفسیر	Lipase	آزمایش	BS-800 اتوآنالایزر	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	67	SD	2.21	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	66.75	SD	1.55	CV	2.32			

