

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High - CV : 0.81	
Mean	112.53	Mean	120.65
SD	0.99	SD	0.98
CV	0.88	CV	0.81
Day-Run	Value	Day-Run	Value
2 - 1	111.6	2 - 1	120
3 - 1	111.4	3 - 1	120.1
4 - 1	111.8	4 - 1	120.5
5 - 1	109.4	5 - 1	119
6 - 1	113	6 - 1	121
9 - 1	113.9	9 - 1	122
11 - 1	113.7	11 - 1	121.8
12 - 1	110.8	12 - 1	120.8

Westgard	تفسیر	MCV	آزمایش	NIHON KOHDEN سل کانتر	دستگاه	10	ماه	1402	سال
Normal Control		Mean	111	SD	1.45	Exp. Date	1402/12/29	Lot. No	

