

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	11.9	
2 - 1	11.3	
3 - 1	11.8	
4 - 1	12	
5 - 1	11.5	
6 - 1	11.56	
7 - 1	11.99	
8 - 1	11.5	
9 - 1	11.6	
12 - 1	11.3	
15 - 1	11.9	
18 - 1	12.3	
21 - 1	11.5	
24 - 1	11.1	
28 - 1	11.5	
30 - 1	11.6	

Westgard	تفسیر	Pb	آزمایش	Atomic Absorption Spectrometry	دستگاه	11	ماه	1402	سال
Control	آزمایشگاه	Mean	11.82	SD	0.36	Exp. Date	1403/1/30	Lot. No	
Next Month	Mean	11.65	SD	0.31	CV	2.68			

