

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	3.43
SD	0.12
CV	3.39
Day-Run	Value
1 - 1	3.3
2 - 1	3.3
3 - 1	3.4
4 - 1	3.5
7 - 1	3.5
8 - 1	3.4
9 - 1	3.3
10 - 1	3.3
11 - 1	3.5
12 - 1	3.6
14 - 1	3.3
15 - 1	3.5
16 - 1	3.6
17 - 1	3.4
18 - 1	3.4
21 - 1	3.3
23 - 1	3.3
24 - 1	3.5
25 - 1	3.3
28 - 1	3.5
29 - 1	3.6
30 - 1	3.6

High - CV : 2.09	
Mean	8.13
SD	0.17
CV	2.09
Day-Run	Value
1 - 1	8
2 - 1	8.2
3 - 1	8
4 - 1	8.1
7 - 1	8.2
8 - 1	8.1
9 - 1	8.5
10 - 1	8.5
11 - 1	8.1
12 - 1	8.4
14 - 1	8
15 - 1	8.1
16 - 1	7.9
17 - 1	8.2
18 - 1	8
21 - 1	8
23 - 1	8
24 - 1	8.2
25 - 1	8.3
28 - 1	8
29 - 1	8
30 - 1	8

Westgard	تفسیر	Phosphorus ( P )	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	3.47	SD	0.1	Exp. Date	1403/1/17	Lot. No	035

