

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	4.8	
3 - 1	4.79	
4 - 1	4.6	1:2S
5 - 1	4.82	
6 - 1	4.6	1:2S
7 - 1	4.8	
9 - 1	4.85	
10 - 1	4.7	
11 - 1	4.87	
12 - 1	4.83	
13 - 1	4.7	
14 - 1	4.8	
16 - 1	4.75	
17 - 1	4.8	
18 - 1	4.81	
19 - 1	4.76	1:2S
20 - 1	4.69	
21 - 1	4.8	
23 - 1	4.66	
24 - 1	4.7	
25 - 1	4.88	
26 - 1	4.65	
27 - 1	4.8	
28 - 1	4.72	
30 - 1	4.8	

Westgard	تفسیر	Potassium (K)	آزمایش	AUDICOM	دستگاه	10	ماه	1402	سال
Control	زیست شیمی	Mean	4.75	SD	0.07	Exp. Date	1404/1/14	Lot. No	210202-17
Next Month	Mean	4.76	SD	0.08	CV	1.65			

