

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	7
SD	0.09
CV	1.24
Day-Run	Value
2 - 1	7.1
3 - 1	7.2
4 - 1	7.1
5 - 1	7
6 - 1	7
7 - 1	7
9 - 1	7
10 - 1	6.9
11 - 1	7
12 - 1	7
13 - 1	7
14 - 1	7
16 - 1	7
17 - 1	7
18 - 1	7.1
19 - 1	7
20 - 1	7
21 - 1	7
23 - 1	6.8
24 - 1	6.9
25 - 1	7
26 - 1	7.1
27 - 1	6.8
28 - 1	7
30 - 1	7

High - CV : 1.27	
Mean	8.33
SD	0.11
CV	1.27
Day-Run	Value
2 - 1	8.4
3 - 1	8.5
4 - 1	8.2
5 - 1	8.3
6 - 1	8.4
7 - 1	8.3
9 - 1	8.4
10 - 1	8.4
11 - 1	8.3
12 - 1	8.1
13 - 1	8.3
14 - 1	8.1
16 - 1	8.3
17 - 1	8.3
18 - 1	8.4
19 - 1	8.4
20 - 1	8.4
21 - 1	8.4
23 - 1	8.3
24 - 1	8.3
25 - 1	8.4
26 - 1	8.2
27 - 1	8.2
28 - 1	8.5
30 - 1	8.4

Westgard	تفسیر	Protein Total	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	7	SD	0.13	Exp. Date	1403/1/17	Lot. No	035

