

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

| Normal | |
|---------|-------|
| Mean | 4.41 |
| SD | 0.03 |
| CV | 0.64 |
| Day-Run | Value |
| 2 - 1 | 4.42 |
| 3 - 1 | 4.44 |
| 6 - 1 | 4.4 |
| 9 - 1 | 4.37 |
| 11 - 1 | 4.38 |
| 12 - 1 | 4.39 |
| 19 - 1 | 4.44 |

| High - CV : 1.1 | |
|-----------------|-------|
| Mean | 4.9 |
| SD | 0.05 |
| CV | 1.1 |
| Day-Run | Value |
| 2 - 1 | 4.8 |
| 4 - 1 | 4.91 |
| 6 - 1 | 4.95 |
| 9 - 1 | 4.9 |
| 11 - 1 | 4.91 |
| 12 - 1 | 4.89 |
| 19 - 1 | 4.97 |

| | | | | | | | | | |
|----------------|-------|------|--------|-----------------------|--------|-----------|------------|---------|-----|
| Westgard | تفسیر | RBC | آزمایش | NIHON KOHDEN سل کانتر | دستگاه | 10 | ماه | 1402 | سال |
| Normal Control | R&D | Mean | 4.41 | SD | 0.07 | Exp. Date | 1402/12/29 | Lot. No | |

