

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	110	
2 - 1	109.1	
3 - 1	110.3	
4 - 1	104.6	
7 - 1	109.5	
8 - 1	110.1	
8 - 2	105.6	
10 - 1	109.7	
12 - 1	110.6	
14 - 1	108.7	
15 - 1	104.7	
16 - 1	110.3	
17 - 1	106.5	
18 - 1	107.6	
21 - 1	109.2	
23 - 1	105.3	
24 - 1	104.6	
25 - 1	110.2	
28 - 1	108.6	
29 - 1	107.6	
30 - 1	105.6	

Westgard	تفسیر	RF	آزمایش	BS-480 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	aptec	Mean	106.88	SD	2.05	Exp. Date	1403/8/19	Lot. No	06095
Next Month	Mean	108.02	SD	2.17	CV	2.01			

