

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	99.7	
2 - 1	98.7	
3 - 1	98.9	
4 - 1	99.2	
5 - 1	99.8	
7 - 1	99.4	
8 - 1	99.2	
9 - 1	99.5	
11 - 1	99.3	
14 - 1	99.6	
16 - 1	99.8	
18 - 1	99.1	
21 - 1	98.9	
24 - 1	99.9	
26 - 1	99.2	
28 - 1	98.6	
29 - 1	99.3	
30 - 1	99.4	

Westgard	تفسیر	Selenium	آزمایش	Atomic Absorption Spectrometry	دستگاه	11	ماه	1402	سال
Control	آزمایشگاه	Mean	99.13	SD	0.48	Exp. Date	1403/1/30	Lot. No	
Next Month	Mean	99.31	SD	0.38	CV	0.38			

