

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	148	
3 - 1	147.9	
4 - 1	147.5	
5 - 1	147	
6 - 1	148.1	
7 - 1	147.3	
9 - 1	147.5	
10 - 1	148.6	
11 - 1	146.9	
12 - 1	147	
13 - 1	148.5	
14 - 1	147	
16 - 1	148	
17 - 1	146.9	
18 - 1	147	
19 - 1	148.2	
20 - 1	147.1	
21 - 1	148.1	
23 - 1	147.1	
24 - 1	148.2	
25 - 1	148.1	
26 - 1	147.6	
27 - 1	148.2	
28 - 1	147.1	
30 - 1	148.2	

Westgard	تفسیر	Sodium (Na)	آزمایش	AUDICOM	دستگاه	10	ماه	1402	سال
Control	زیست شیمی	Mean	147.72	SD	0.52	Exp. Date	1404/1/14	Lot. No	2103061-4
Next Month	Mean	147.64	SD	0.56	CV	0.38			

