

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal	
Mean	1.02
SD	0.06
CV	6.35
Day-Run	Value
2 - 1	1
3 - 1	1.05
4 - 1	0.96
5 - 1	1.05
6 - 1	0.9
7 - 1	1.05
9 - 1	1.1
10 - 1	1.1
11 - 1	1.05
12 - 1	0.93
13 - 1	1
14 - 1	1.05
16 - 1	1
17 - 1	0.9
18 - 1	1.1
19 - 1	0.9
20 - 1	1.04
21 - 1	1.12
23 - 1	1
24 - 1	1.11
25 - 1	1
26 - 1	1.03
27 - 1	1.04
28 - 1	1.05
30 - 1	1

High - CV : 2.71	
Mean	4.26
SD	0.12
CV	2.71
Day-Run	Value
2 - 1	4.25
3 - 1	4.3
4 - 1	4.4
5 - 1	4.1
6 - 1	4.36
7 - 1	4.25
9 - 1	4.15
10 - 1	4.3
11 - 1	4.3
12 - 1	4.15
13 - 1	4.2
14 - 1	4.12
16 - 1	4.5
17 - 1	4.25
18 - 1	4.2
19 - 1	4.5
20 - 1	4.29
21 - 1	4.1
23 - 1	4.2
24 - 1	4.37
25 - 1	4.1
26 - 1	4.3
27 - 1	4.4
28 - 1	4.3
30 - 1	4.2

Westgard	تفسیر	Bili T	آزمایش	اتوآنالایزر BS-800	دستگاه	10	ماه	1402	سال
Normal Control	biorex	Mean	1.01	SD	0.07	Exp. Date	1403/1/17	Lot. No	035

