

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	12.2	
2 - 1	12.5	
3 - 1	11.6	
4 - 1	9.9	1:2S
7 - 1	10.35	1:2S 2:2S Cusum Reject
8 - 1	11.7	
9 - 1	11.5	
10 - 1	11.7	
11 - 1	11	
12 - 1	11.7	
14 - 1	11.75	
15 - 1	11.5	
16 - 1	12	
17 - 1	11.9	
18 - 1	11.3	
21 - 1	12.2	
23 - 1	12.7	
24 - 1	11.5	
25 - 1	11.6	
26 - 1	11	
28 - 1	10.7	
29 - 1	10.5	
30 - 1	11.7	

Westgard	تفسیر	T4	آزمایش	SIEMENS Centaur CP	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	11.7	SD	0.65	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	11.5	SD	0.68	CV	5.9			

