

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	12.7	
3 - 1	12.2	
4 - 1	10.7	
5 - 1	11.5	
6 - 1	11.9	
7 - 1	12.2	
9 - 1	10.9	
10 - 1	11.8	
11 - 1	11.9	
12 - 1	11.6	
13 - 1	12.2	
14 - 1	12.6	
16 - 1	12.1	
17 - 1	12.7	
18 - 1	11.7	
19 - 1	10.6	
20 - 1	11.3	
21 - 1	12	
23 - 1	11.6	
24 - 1	10.9	
25 - 1	10.5	1:2S
26 - 1	10.9	
27 - 1	12	
28 - 1	12.3	
30 - 1	11.8	

Westgard	تفسیر	T4	آزمایش	SIEMENS Centaur CP	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	11.84	SD	0.62	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	11.7	SD	0.65	CV	5.56			

