

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	173
SD	2.45
CV	1.42
Day-Run	Value
1 - 1	171
2 - 1	176
3 - 1	170
4 - 1	175
7 - 1	170
8 - 1	170
9 - 1	173
10 - 1	175
11 - 1	170
12 - 1	170
14 - 1	177
15 - 1	170
16 - 1	175
17 - 1	174
18 - 1	173
21 - 1	171
23 - 1	173
24 - 1	176
25 - 1	173
28 - 1	174
29 - 1	173
30 - 1	177

High - CV : 1.36	
Mean	230
SD	3.13
CV	1.36
Day-Run	Value
1 - 1	233
2 - 1	234
3 - 1	236
4 - 1	232
7 - 1	232
8 - 1	225
9 - 1	232
10 - 1	228
11 - 1	230
12 - 1	228
14 - 1	232
15 - 1	227
16 - 1	228
17 - 1	231
18 - 1	230
21 - 1	225
23 - 1	230
24 - 1	225
25 - 1	233
28 - 1	225
29 - 1	230
30 - 1	232
30 - 2	232

Westgard	تفسیر	Triglycerides	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	172.01	SD	2.57	Exp. Date	1403/1/17	Lot. No	035

