

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	307.6	
2 - 1	320.7	
3 - 1	300	
4 - 1	307.9	
7 - 1	322.8	1:2S
8 - 1	310	
9 - 1	319.6	
10 - 1	320	
11 - 1	323	1:2S
12 - 1	308	
14 - 1	322	1:2S
15 - 1	320	
16 - 1	308	
17 - 1	309	
18 - 1	307.9	
21 - 1	313	
23 - 1	311.3	
24 - 1	314.3	
25 - 1	303.9	
28 - 1	304.9	
29 - 1	307.7	
30 - 1	303.1	

Westgard	تفسیر	TIBC	آزمایش	BS-800 اتوآنالایزر	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	310.17	SD	6.5	Exp. Date	1403/1/17	Lot. No	035
Next Month	Mean	312.03	SD	7.11	CV	2.28			

