

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	3.82	
2 - 1	3.95	
3 - 1	3.85	
4 - 1	4.02	
7 - 1	3.6	
8 - 1	3.68	
9 - 1	3.53	
10 - 1	3.8	
11 - 1	3.95	
12 - 1	3.9	
14 - 1	3.72	
15 - 1	4.1	
16 - 1	3.8	
17 - 1	3.4	
18 - 1	3.56	
21 - 1	3.84	
23 - 1	3.97	
24 - 1	3.79	
25 - 1	4.1	
26 - 1	3.6	
28 - 1	3.82	
29 - 1	3.95	
30 - 1	3.86	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	11	ماه	1402	سال
Control	biorex	Mean	3.75	SD	0.25	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	3.81	SD	0.18	CV	4.79			

