

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	3.51	
3 - 1	3.76	
4 - 1	3.22	
5 - 1	3.58	
6 - 1	3.92	
7 - 1	4.05	
9 - 1	3.86	
10 - 1	3.48	
11 - 1	3.6	
12 - 1	3.65	
13 - 1	4.02	
14 - 1	3.85	
16 - 1	3.65	
17 - 1	3.8	
18 - 1	3.79	
19 - 1	4.1	
20 - 1	3.95	
21 - 1	4	
23 - 1	3.64	
24 - 1	3.25	
25 - 1	3.4	
26 - 1	3.9	
27 - 1	4.05	
28 - 1	3.8	
30 - 1	3.95	

Westgard	تفسیر	TSH	آزمایش	SIEMENS Centaur CP	دستگاه	10	ماه	1402	سال
Control	biorex	Mean	3.71	SD	0.27	Exp. Date	1403/6/19	Lot. No	36-02-028
Next Month	Mean	3.75	SD	0.25	CV	6.59			

