

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
2 - 1	495	
3 - 1	536	
4 - 1	555	
5 - 1	511	
6 - 1	516	
7 - 1	501	
9 - 1	489	
10 - 1	475	
11 - 1	482	
12 - 1	527	
13 - 1	538	
14 - 1	510	
16 - 1	495	
17 - 1	486	
18 - 1	535	
19 - 1	554	
20 - 1	521	
21 - 1	508	
23 - 1	534	
24 - 1	495	
25 - 1	482	
26 - 1	515	
27 - 1	520	
28 - 1	500	
30 - 1	529	

Westgard	تفسیر	Testosterone	آزمایش	2000 ایمولایت XPI-	دستگاه	10	ماه	1402	سال
Control	Labex	Mean	501.63	SD	55.23	Exp. Date	1403/8/30	Lot. No	36502028
Next Month	Mean	512.36	SD	22.44	CV	4.38			

