

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Levey Jennings Chart

Normal	
Mean	4.78
SD	0.14
CV	3
Day-Run	Value
1 - 1	4.8
2 - 1	4.5
3 - 1	4.7
4 - 1	4.6
7 - 1	4.7
8 - 1	4.6
9 - 1	4.6
10 - 1	4.8
11 - 1	4.7
12 - 1	4.8
14 - 1	4.7
15 - 1	4.7
16 - 1	4.8
17 - 1	4.8
18 - 1	4.8
21 - 1	4.9
23 - 1	4.9
24 - 1	5
25 - 1	4.9
28 - 1	4.9
29 - 1	5.1
30 - 1	4.9

High - CV : 2.38	
Mean	8.53
SD	0.2
CV	2.38
Day-Run	Value
1 - 1	8.6
2 - 1	8.4
3 - 1	8.6
4 - 1	8.8
7 - 1	8.5
8 - 1	8.6
9 - 1	8.6
10 - 1	8.5
11 - 1	8
12 - 1	8.6
14 - 1	8
15 - 1	8.5
16 - 1	8.5
17 - 1	8.6
18 - 1	8.6
21 - 1	8.5
23 - 1	8.6
24 - 1	8.4
25 - 1	8.6
28 - 1	8.6
29 - 1	8.8
30 - 1	8.8

Westgard	تفسیر	Uric Acid	آزمایش	اتوآنالایزر BS-800	دستگاه	11	ماه	1402	سال
Normal Control	biorex	Mean	4.8	SD	0.25	Exp. Date	1403/1/17	Lot. No	035

