

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
19 - 1	584	
20 - 1	580	
21 - 1	559	
23 - 1	550	
24 - 1	563	
25 - 1	578	
26 - 1	583	
27 - 1	590	
28 - 1	568	
30 - 1	581	
1 - 1	557	

Westgard	تفسیر	Vitamin B12(Cobalamin)	آزمایش	Maglumi	دستگاه	10	ماه	1402	سال
Control		Maglumi	Mean	571.2	SD	10.97	Exp. Date	1403/10/18	Lot. No
Next Month		Mean	572.09	SD	13.21	CV	2.31		

