

آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	28.9	
2 - 1	29.1	
3 - 1	29.4	
4 - 1	29.4	
7 - 1	28.8	
8 - 1	28.6	
9 - 1	29.6	
10 - 1	29.8	
11 - 1	29.6	
12 - 1	28.8	
14 - 1	29.6	
15 - 1	28.9	
17 - 1	29.7	
21 - 1	29.8	
23 - 1	29.3	
24 - 1	28.5	
25 - 1	29.7	
26 - 1	28.9	
28 - 1	28.2	
29 - 1	28.6	
30 - 1	29.4	

Westgard	تفسیر	V D3	آزمایش	HPLC Agilent 1100 -1	دستگاه	11	ماه	1402	سال
Control	آزمایشگاه	Mean	29.14	SD	0.67	Exp. Date	1403/1/17	Lot. No	
Next Month	Mean	29.17	SD	0.48	CV	1.65			

