

# آزمایشگاه پاتوبیولوژی و ژنتیک کاوش گرگان

## Control Serum - Levey Jennings Chart

Normal		High		Low	
Mean	7.53	Mean	19.99	Mean	3.01
SD	0.24	SD	0.59	SD	0.11
CV	3.14	CV	2.94	CV	3.74
Day-Run	Value	Day-Run	Value	Day-Run	Value
2 - 1	7.7	2 - 1	21.1	2 - 1	3
3 - 1	7.3	3 - 1	20	3 - 1	3
6 - 1	7.1	6 - 1	19.8	3 - 2	2.8
9 - 1	7.7	9 - 1	20	6 - 1	3
11 - 1	7.6	11 - 1	19.1	9 - 1	3.1
12 - 1	7.7	12 - 1	19.9	11 - 1	3
19 - 1	7.6	19 - 1	20	12 - 1	2.81
				19 - 1	3.2

Westgard	تفسیر	WBC		آزمایش	NIHON KOHDEN سل کانتتر	دستگاه	10	ماه	1402	سال
Normal Control	R&D	Mean	7.2	SD	0.28	Exp. Date	1402/12/29		Lot. No	
High Control	R&D	Mean	19.3	SD	0.62	Exp. Date	1402/12/29		Lot. No	
Low Control	R&D	Mean	2.9	SD	0.13	Exp. Date	1402/12/29		Lot. No	

