

آزمایشگاه پاتوبیولوژی و ژنتیک گرگان

مرکز فنون آزمایشگاهی خاورمیانه

ME101

کد آزمون:

Clinical Chemistry

Clinical chemistry PT/EQA, 1403, 2nd round

نام آزمون:

عنوان تقویم:

آنالیت های کمی

Performance	CV-B	D%-B	SDI-B	AV-B	نتیجه B	CV-A	D%-A	SDI-A	AV-A	نتیجه A	واحد	Peer Group
✓	8.68	4.84	0.63	9.44	9.9	9.94	3.70	0.48	4.92	5.1	mg/dl	TOO-Bior
✓	7.28	-0.22	-0.03	6.70	6.7	6.21	1.09	0.15	3.75	3.8	g/dl	BCG-B
✓	13.93	-0.42	-0.03	87.28	87	14.40	16.52	1.07	168.77	197	U/l	DEA-B
✓						4.80	1.43	0.90	149.90	152	meq/l	ISE Dir Audio
✓						7.68	7.75	0.87	0.88	0.95	mg/dl	Jaffe-B
✓	17.13	2.32	0.15	448.25	454	9.62	-4.66	-0.31	291.59	277	µg/dl	Dire colorim Pishta
✓	11.12	-5.00	-0.88	43.52	42	9.69	1.87	0.33	27.43	28.2	mg/dl	UV-Bi
✓	4.83	3.60	0.47	128.27	133	4.74	-1.10	-0.14	84.87	84	mg/dl	GOD-H Bior
✓	5.91	8.28	2.12	11.41	12.4	5.02	1.68	0.43	5.68	5.8	g/dl	Biore
✓	28.54	-8.85	-0.74	51.56	47	23.72	4.63	0.39	33.45	35	mg/dl	Othe
✓	7.01	7.72	1.02	282.20	304	6.82	-1.88	-0.25	134.53	132	mg/dl	CHO PAP-B
✓	11.10	15.05	1.31	120.97	140	8.59	14.04	1.22	61.56	70.5	U/l	CNPO Bior
✓	27.15	-25.99	-1.35	0.92	0.7	21.16	-15.60	-0.81	0.47	0.4	mg/dl	Jendra Grof-B
✓	16.91	9.31	0.62	178.39	195	15.88	1.67	0.11	76.72	78	mg/dl	Dire metho Other
✓	35.14	-21.08	-1.10	0.31	0.24	31.31	-28.18	-1.47	0.15	0.11	mg/dl	Jendra Grof-B

Performance	CV-B	D%-B	SDI-B	AV-B	نتیجه B	CV-A	D%-A	SDI-A	AV-A	نتیجه A	واحد	Peer Group
?	15.14	21.75	2.72	500.72	612	13.52	16.79	2.10	278.79	328	U/l	IFCC_E
✓	11.39	0.80	0.05	152.53	153	9.56	-6.87	-0.46	101.37	94	µg/dl	Ferri Bion
✓	13.30	14.66	1.47	344.55	398	9.46	4.06	0.41	91.30	95	U/l	IFCC-E
✓	15.77	321.10	17.84	53.77	228	21.96	1.52	0.08	31.02	32	U/l	Colorin Bion
✓	7.19	2.13	0.28	366.21	374	5.39	4.83	0.63	135.65	142	mg/dl	GPO-F Del
✓	13.01	0.12	0.01	18.87	19	9.78	16.99	1.36	22.17	26	U/l	IFCC- PLP-D
✓						10.56	1.34	0.08	28.54	29	U/l	IFCC- PLP-D
✓	5.61	5.21	1.30	15.40	16.2	6.55	-3.75	-0.94	8.10	7.8	mg/dl	Arsena ASAL
✓	10.18	8.37	1.07	5.26	5.7	3.96	1.56	0.20	2.56	2.6	mg/dl	Molyb UV ASAL